

Czech Metrology Institute Notified Body No. 1383

V 3112

Okružní 31, 638 00 Brno, Czech Republic tel. +420 545 555 111, fax +420 545 222 728 www.cmi.cz

EU-TYPE EXAMINATION CERTIFICATE

Number: TCM 221/24 - 5969

Addition 2

This addition replaces all previous versions of this certificate in full wording.

Page 1 from 7 pages

In accordance:

with Directive 2014/32/EU of the European Parliament and of the Council on the harmonisation of the laws of the Member States relating to the making available on the market of measuring instruments (implemented in Czech Republic by Government Order No. 120/2016 Coll.).

Manufacturer:

Elexim, a.s.

Riegrovo náměstí 179/14

767 01 Kroměříž Czech Republic

For:

active electrical energy meter - 3-phase

type: E²MeVOLT accuracy class: B

mechanical environment class: M1 electromagnetic environment class: E2

temperature range: -10 °C...+55 °C

Valid until:

21 November 2034

Document No:

0511-CS-A022-24

Description:

Essential characteristics, approved conditions and special conditions, if any, are described in

this certificate.

Date of issue:

23 June 2025

Certificate approved by:

RNDr. Pavel Klenovský

Cesky metitut 4ntisem: Antisem: Antise